

**PRODUCT INFORMATION**

<b>Product SKU</b>	ANARC
<b>Product Name</b>	Anaplasma Run Control
<b>Kit Lot Number</b>	20071701
<b>Source Material</b>	<i>Anaplasma phagocytophilum</i>
<b>Date of Expiration</b>	2021-12-31
<b>Base Matrix</b>	Synthetic Matrix
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House Anaplasma Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
20071701	Anaplasma Run Control	10,000 cp/mL	Pass	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

<b>Quality Assurance</b>	<b>Date</b>
	13 Nov 20