

PRODUCT INFORMATION

Product SKU	CMVL101
Product Name	CMV Low Run Control
Kit Lot Number	20021302
Source Material	Inactivated CMV
Date of Expiration	2023-01-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Traceability	1 st WHO International Standard (09/162)
Testing Platform	In-House CMV Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Testing Results	Quantity
20021302	CMV Low Run Control	Pass (Comparison of target value against reference standard)	5 vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
<i>Saylor Dupli</i>	02 Mar 20

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com

CMVL101 LN20021302 Rev 01.00