

**PRODUCT INFORMATION**

<b>Product SKU</b>	CMVL201
<b>Product Name</b>	CMV 1E3 IU/mL Precision
<b>Kit Lot Number</b>	18032708
<b>Source Material</b>	Inactivated CMV
<b>Date of Expiration</b>	2020-03
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservatives</b>	ProClin® 300
<b>Fill Volume</b>	1.4 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Traceability</b>	1 <sup>st</sup> WHO International Standard (09/162)
<b>Testing Platform</b>	In-House CMV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results	Quantity
18032708	CMV 1E3 IU/mL Precision	1,000 IU/mL	Pass (Comparison of target value against reference standard)	15 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	21 Sep 18