

PRODUCT INFORMATION

Product SKU	CMVL201
Product Name	CMV 1E3 IU/mL Precision
Kit Lot Number	21051112
Source Material	Inactivated CMV
Date of Expiration	2023-04-30
Base Matrix	True EDTA Plasma
Preservatives	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	1 st WHO International Standard (09/162)
Testing Platform	In-House CMV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results	Quantity
21051112	CMV 1E3 IU/mL Precision	1,000 IU/mL	Pass (Comparison of target value against reference standard)	9 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	25 May 21