

**PRODUCT INFORMATION**

<b>Product SKU</b>	CMVP100
<b>Product Name</b>	CMV Verification Panel
<b>Kit Lot Number</b>	18092009
<b>Source Material</b>	Inactivated CMV
<b>Date of Expiration</b>	2021-09-30
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Traceability</b>	1 <sup>st</sup> WHO International Standard (09/162)
<b>Testing Platform</b>	In-House CMV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.


**TESTING DATA**

Lot Number	Product Description	Target Concentrations*	Results
18092004	CMV Panel Member 1: 4E2 IU/mL	400 IU/mL	Pass
18092005	CMV Panel Member 2: 4E3 IU/mL	4,000 IU/mL	Pass
18092006	CMV Panel Member 3: 4E4 IU/mL	40,000 IU/mL	Pass
18092007	CMV Panel Member 4: 4E5 IU/mL	400,000 IU/mL	Pass
18092008	CMV Panel Member 5: 4E6 IU/mL	4,000,000 IU/mL	Pass

\* 1 International Unit = 1 digital PCR copy

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	24 Oct 18