

**PRODUCT INFORMATION**

Product SKU	CMVP200
Product Name	CMV Verification Panel 1.4mL
Kit Lot Number	21042810
Source Material	Inactivated CMV
Date of Expiration	2023-04-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	1 <sup>st</sup> WHO International Standard (09/162)
Testing Platform	In-House CMV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
21042811	CMV Panel Member 1: 2E2 IU/mL	200 IU/mL	Pass
21042812	CMV Panel Member 2: 4E2 IU/mL	400 IU/mL	Pass
21042813	CMV Panel Member 3: 4E3 IU/mL	4,000 IU/mL	Pass
21042814	CMV Panel Member 4: 4E4 IU/mL	40,000 IU/mL	Pass
21042815	CMV Panel Member 5: 4E5 IU/mL	400,000 IU/mL	Pass
21042816	CMV Panel Member 6: 4E6 IU/mL	4,000,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	12 May 21