

PRODUCT INFORMATION

Product SKU	COV000
Product Name	SARS-CoV-2 Negative
Lot Number	20033002
Date of Manufacture	2020-04-02
Date of Expiration	2021-03-31
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	For Research Use Only. Not for use in diagnostic procedures.
Testing Platform	In-House RNase P Assay

This standard material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
20033002	SARS-CoV-2 Negative	Human gDNA	75,000 cp/mL	Pass	5 Vials

APPROVAL

Quality Assurance	Date
	16 Nov 20