

PRODUCT INFORMATION

Product SKU	COV000CE
Product Name	SARS-CoV-2 Negative Run Control
Lot Number	20052601
Date of Manufacture	2020-06-03
Date of Expiration	2021-11-30
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House RNase P Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Target Concentration	Testing Results	Quantity
20052601	SARS-CoV-2 Negative Run Control	Human gDNA	75,000 cp/mL RNase P DNA	Pass 154,000 cp/mL RNase P (DNA + RNA)	5 Vials

APPROVAL

Quality Assurance	Date
	07 Jun 21