

PRODUCT INFORMATION

Product SKU	COV019CE
Product Name	SARS-CoV-2 Positive Run Control
Lot Number	21021102
Date of Manufacture	2021-07-01
Date of Expiration	2022-12-31
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House E, N, S, ORF1ab, RdRP, RNase P Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Target Concentrations	Testing Results	Quantity
21021102	SARS-CoV-2 Positive Run Control	Envelope (E) gene RNA Transcript	200,000 cp/mL	Pass	5 Vials
		Nucleocapsid (N) gene RNA Transcript	200,000 cp/mL	Pass	
		Spike Protein (S) gene RNA Transcript	200,000 cp/mL	Pass	
		ORF1ab gene RNA Transcript	200,000 cp/mL	Pass	
		RdRP RNA Transcript	200,000 cp/mL	Pass	
		Human gDNA	75,000 cp/mL RNase P DNA	Pass 154,000 cp/mL RNase P (DNA + RNA)	

APPROVAL

Quality Assurance 	Date 12 Jun 21
---	--------------------------