

PRODUCT INFORMATION

Product SKU	COVFLU
Product Name	SARS-CoV-2, Flu, RSV Positive Run Control
Lot Number	21010704
Date of Expiration	2022-12-31
Base Matrix	Stabilizing Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House Influenza A H1N1, Influenza B, RSV A, N Gene, RNase P Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Target Concentrations	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21010704	SARS-CoV-2, Flu, RSV Positive Run Control	Influenza A H1N1	10,000 cp/mL	Pass	5 Vials
		Influenza B	10,000 cp/mL	Pass	
		RSV A	10,000 cp/mL	Pass	
		SARS-CoV-2	10,000 cp/mL	Pass	
		Human gDNA	10,000 cp/mL RNase P DNA	Pass	

APPROVAL

Quality Assurance	Date
	05 Feb 21