

**PRODUCT INFORMATION**

<b>Product SKU</b>	CTNGP200
<b>Product Name</b>	CT/NG Verification Panel
<b>Kit Lot Number</b>	21121611
<b>Source Material 1</b>	<i>Chlamydia trachomatis</i>
<b>Source Material 2</b>	<i>Neisseria gonorrhoeae</i>
<b>Date of Expiration</b>	2023-06-30
<b>Matrix</b>	Multiprep Specimen Diluent
<b>Preservative</b>	Sodium Azide
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Testing Platform</b>	cobas® 4800 CT/NG Test

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
21121602	CT 3x LOD	CT: 600 cp/mL	Pass
21121603	CT 5x LOD	CT: 1,000 cp/mL	Pass
21121604	CT 10x LOD	CT: 2,000 cp/mL	Pass
21121605	NG 3x LOD	NG: 3,000 cp/mL	Pass
21121606	NG 5x LOD	NG: 5,000 cp/mL	Pass
21121607	NG 10x LOD	NG: 10,000 cp/mL	Pass
21121608	CT/NG 3x LOD	CT: 600 cp/mL	Pass
		NG: 3,000 cp/mL	
21121609	CT/NG 5x LOD	CT: 1,000 cp/mL	Pass
		NG: 5,000 cp/mL	
21121610	CT/NG 10x LOD	CT: 2,000 cp/mL	Pass
		NG: 10,000 cp/mL	

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

<b>Quality Assurance</b> 	<b>Date</b> 12 Apr 22
---	--------------------------