

**PRODUCT INFORMATION**

<b>Product SKU</b>	CTNGP200
<b>Product Name</b>	CT/NG Verification Panel
<b>Kit Lot Number</b>	21080310
<b>Source Material 1</b>	<i>Chlamydia trachomatis</i>
<b>Source Material 2</b>	<i>Neisseria gonorrhoeae</i>
<b>Date of Expiration</b>	2023-01-31
<b>Matrix</b>	Multiprep Specimen Diluent
<b>Preservative</b>	Sodium Azide
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Testing Platform</b>	cobas® 4800 CT/NG Test

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
21080301	CT 3x LOD	CT: 600 cp/mL	Pass
21080302	CT 5x LOD	CT: 1,000 cp/mL	Pass
21080303	CT 10x LOD	CT: 2,000 cp/mL	Pass
21080304	NG 3x LOD	NG: 3,000 cp/mL	Pass
21080305	NG 5x LOD	NG: 5,000 cp/mL	Pass
21080306	NG 10x LOD	NG: 10,000 cp/mL	Pass
21080307	CT/NG 3x LOD	CT: 600 cp/mL NG: 3,000 cp/mL	Pass
21080308	CT/NG 5x LOD	CT: 1,000 cp/mL NG: 5,000 cp/mL	Pass
21080309	CT/NG 10x LOD	CT: 2,000 cp/mL NG: 10,000 cp/mL	Pass

<b>Bioburden Testing Results</b>	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	06 Jan 22