

**PRODUCT INFORMATION**

<b>Product SKU</b>	EBVP100
<b>Product Name</b>	EBV Verification Panel
<b>Kit Lot Number</b>	18062509
<b>Source Material</b>	Inactivated EBV
<b>Date of Expiration</b>	2020-06-30
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Traceability</b>	1 <sup>st</sup> WHO International Standard (09/260)
<b>Testing Platform</b>	In-House EBV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.


**TESTING DATA**

Lot Number	Product Description	Target Concentrations (IU/mL)*	Results
18062510	EBV Panel Member 1: 5E2 IU/mL	500 IU/mL	Pass
18062511	EBV Panel Member 2: 5E3 IU/mL	5,000 IU/mL	Pass
18062512	EBV Panel Member 3: 5E4 IU/mL	50,000 IU/mL	Pass
18062513	EBV Panel Member 4: 5E5 IU/mL	500,000 IU/mL	Pass
18062514	EBV Panel Member 5: 5E6 IU/mL	5,000,000 IU/mL	Pass

\*1 International Unit = 0.49 digital PCR copy

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	21Oct18