

PRODUCT INFORMATION

Product SKU	EBVP200
Product Name	EBV Verification Panel 1.4mL
Kit Lot Number	21033101
Source Material	Inactivated EBV
Date of Expiration	2023-03-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	1st WHO International Standard (09/260)
Testing Platform	In-House EBV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
21033102	EBV Panel Member 1: 1E2 IU/mL 1.4mL	100 IU/mL	Pass
21033103	EBV Panel Member 2: 5E2 IU/mL 1.4mL	500 IU/mL	Pass
21033104	EBV Panel Member 3: 5E3 IU/mL 1.4mL	5,000 IU/mL	Pass
21033105	EBV Panel Member 4: 5E4 IU/mL 1.4mL	50,000 IU/mL	Pass
21033106	EBV Panel Member 5: 5E5 IU/mL 1.4mL	500,000 IU/mL	Pass
21033107	EBV Panel Member 6: 5E6 IU/mL 1.4mL	5,000,000 IU/mL	Pass
21033108	EBV Panel Member 7: 5E7 IU/mL 1.4mL	50,000,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	19 Apr 21