

PRODUCT INFORMATION

Product SKU	ERLRC
Product Name	Ehrlichia Run Control
Kit Lot Number	19041904
Source Material	<i>Ehrlichia chaffeensis</i>
Date of Expiration	2020-09-30
Base Matrix	Synthetic Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House Ehrlichia Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Testing Results	Quantity
19041904	Ehrlichia Run Control	Pass (Comparison of target value against reference standard)	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
<i>Taylor Suptic</i>	<i>20 May 19</i>

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com

ERLRC LN19041904 Rev 01.00