

PRODUCT INFORMATION

Product SKU	HBVL201
Product Name	HBV 1E3 IU/mL Precision
Kit Lot Number	21060114
Source Material	Inactivated HBV
Date of Expiration	2023-05-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	3 rd WHO International Standard (10/264)
Testing Platform	In-House HBV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21060114	HBV 1E3 IU/mL Precision	1,000 IU/mL	Pass	9 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	21 Jun 21