

PRODUCT INFORMATION

Product SKU	HCVGTP
Product Name	HCV Genotype Panel
Kit Lot Number	20071007
Source Material	Inactivated HCV
Date of Expiration	2021-12-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	For Research Use Only. Not for use in diagnostic procedures.
Traceability	4 th WHO International Standard (06/102)
Testing Platform	Roche COBAS® AmpliPrep/COBAS® TaqMan® HCV v2.0

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Results
20071001	HCV Genotype 1a	20,000 IU/mL	Pass
20071002	HCV Genotype 1b	20,000 IU/mL	Pass
20071003	HCV Genotype 2	20,000 IU/mL	Pass
20071004	HCV Genotype 3	20,000 IU/mL	Pass
20071005	HCV Genotype 4	20,000 IU/mL	Pass
20071006	HCV Genotype 5	20,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	17 Aug 20