

PRODUCT INFORMATION

Product SKU	HCVGTP
Product Name	HCV Genotype Panel
Kit Lot Number	21092212
Source Material	Inactivated HCV
Date of Expiration	2023-02-28
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	For Research Use Only. Not for use in diagnostic procedures.
Traceability	4 th WHO International Standard (06/102)
Testing Platform	Roche COBAS® AmpliPrep/COBAS® TaqMan® HCV v2.0

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Results
21092206	HCV Genotype 1a	20,000 IU/mL	Pass
21092207	HCV Genotype 1b	20,000 IU/mL	Pass
21092208	HCV Genotype 2	20,000 IU/mL	Pass
21092209	HCV Genotype 3	20,000 IU/mL	Pass
21092210	HCV Genotype 4	20,000 IU/mL	Pass
21092211	HCV Genotype 5	20,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	05Oct21