

**PRODUCT INFORMATION**

<b>Product SKU</b>	HCVH202
<b>Product Name</b>	HCV 1E6 IU/mL Precision
<b>Kit Lot Number</b>	22042605
<b>Source Material</b>	Recombinant HCV
<b>Date of Expiration</b>	2024-04-30
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.4 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Traceability</b>	4 <sup>th</sup> WHO International Standard (06/102)
<b>Testing Platform</b>	In-House HCV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
22042605	HCV 1E6 IU/mL Precision	1,000,000 IU/mL	Pass	9 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

<b>Quality Assurance</b> 	<b>Date</b> 09 May 22
---	--------------------------