

**PRODUCT INFORMATION**

|                            |   |
|----------------------------|---|
| <b>Product SKU</b>         | HCVL101   |
| <b>Product Name</b>        | HCV Low Run Control                                 |
| <b>Kit Lot Number</b>      | 21060903  |
| <b>Source Material</b>     | Recombinant HCV                                     |
| <b>Date of Manufacture</b> | 2021-06-21  |
| <b>Date of Expiration</b>  | 2024-05-31  |
| <b>Base Matrix</b>         | True EDTA Plasma                                    |
| <b>Preservative</b>        | ProClin® 300  |
| <b>Fill Volume</b>         | 1.0 mL  |
| <b>Storage Temperature</b> | -20°C or Below                                      |
| <b>Regulatory Status</b>   | For In Vitro Diagnostic Use                         |
| <b>Traceability</b>        | 4 <sup>th</sup> WHO International Standard (06/102) |
| <b>Testing Platform</b>    | In-House HCV Assay                                  |

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

| Lot Number | Product Description | Target Concentration | Testing Results<br>(Comparison of Target Value<br>Against Reference Standard) | Quantity |
|------------|---------------------|----------------------|---|----------|
| 21060903   | HCV Low Run Control | 200 IU/mL            | Pass  | 5 Vials  |

| Bioburden Testing Results                |      |
|--|------|
| Acceptance Criteria: Less than 10 CFU/mL | Pass |

**APPROVAL**

|   |             |
|---|-------------|
| <b>Quality Assurance</b>  | <b>Date</b> |
|  | 02 Jul 21   |