

PRODUCT INFORMATION

Product SKU	HCVL201
Product Name	HCV 1E3 IU/mL Precision
Kit Lot Number	19080903
Source Material	Inactivated HCV Genotype 1a
Date of Expiration	2021-07-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	4 th WHO International Standard (06/102)
Testing Platform	In-House HCV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results	Quantity
19080903	HCV 1E3 IU/mL Precision	1,000 IU/mL	Pass (Comparison of target value against reference standard)	9 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance 	Date 27 Aug 19
---	--------------------------