

**PRODUCT INFORMATION**

<b>Product SKU</b>	HCVP200
<b>Product Name</b>	HCV Verification Panel 1.4mL
<b>Kit Lot Number</b>	21080331
<b>Source Material</b>	Recombinant HCV
<b>Date of Expiration</b>	2023-07-31
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.4 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Traceability</b>	4 <sup>th</sup> WHO International Standard (06/102)
<b>Testing Platform</b>	In-House HCV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
21080320	HCV Panel Member 1: 1E2 IU/mL 1.4mL	100 IU/mL	Pass
21080321	HCV Panel Member 2: 3E2 IU/mL 1.4mL	300 IU/mL	Pass
21080322	HCV Panel Member 3: 3E3 IU/mL 1.4mL	3,000 IU/mL	Pass
21080323	HCV Panel Member 4: 3E4 IU/mL 1.4mL	30,000 IU/mL	Pass
21080324	HCV Panel Member 5: 3E5 IU/mL 1.4mL	300,000 IU/mL	Pass
21080325	HCV Panel Member 6: 3E6 IU/mL 1.4mL	3,000,000 IU/mL	Pass
21080326	HCV Panel Member 7: 3E7 IU/mL 1.4mL	30,000,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

<b>Quality Assurance</b> 	<b>Date</b> 14 Sep 21
---	--------------------------