

PRODUCT INFORMATION

Product SKU	HIVAMR
Product Name	HIV-1 Analytical Measurement Range Panel
Kit Lot Number	21040705
Source Material	Inactivated HIV-1
Date of Expiration	2024-03-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	3 rd WHO International Standard (10/152)
Testing Platform	In-House HIV-1 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
21040702	HIV-1 AMR Panel Member 1: 6E2 cp/mL	600 cp/mL	Pass
21040703	HIV-1 AMR Panel Member 2: 6E4 cp/mL	60,000 cp/mL	Pass
21040704	HIV-1 AMR Panel Member 3: 6E6 cp/mL	6,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	20Apr24