

PRODUCT INFORMATION

Product SKU	HIVH102
Product Name	HIV-1 High Run Control
Kit Lot Number	16112805
Source Material	Inactivated HIV Virus
Date of Expiration	2018-06
Base Matrix	True EDTA Plasma
Preservatives	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or below
Intended Use	For In Vitro Diagnostic Use
Traceability	3 rd HIV-1 International Standard (10/152)
Testing Platform	In-house HIV-1 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Testing Results	Quantity
16112805	HIV-1 High Run Control	Pass (Comparison of target value against reference standard)	5 vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	19 Dec 16

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com

HIVH102 LN16112805 Rev 01.00