

PRODUCT INFORMATION

Product SKU	HIVH202
Product Name	HIV-1 1E5 cp/mL Precision
Kit Lot Number	21042902
Source Material	Inactivated HIV
Date of Expiration	2023-04-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.4 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostic Use
Traceability	3 rd WHO International Standard (10/152)
Testing Platform	In-House HIV-1 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21042902	HIV-1 1E5 cp/mL Precision	100,000 cp/mL	Pass	9 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	11 May 21