

**PRODUCT INFORMATION**

<b>Product SKU</b>	HPV16C
<b>Product Name</b>	HPV Genotype 16 Control
<b>Kit Lot Number</b>	20013102
<b>Source Material</b>	HPV Genotype 16 Cell Line
<b>Date of Expiration</b>	2023-01-31
<b>Base Matrix</b>	PreservCyt®
<b>Fill Volume</b>	5.0 mL
<b>Storage Temperature</b>	2-8°C
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House HPV-16 Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

<b>Lot Number</b>	<b>Product Description</b>	<b>Target Concentration</b>	<b>Testing Results</b> (Comparison of Target Value Against Reference Standard)	<b>Quantity</b>
20013102	HPV Genotype 16 Control	HPV-16: 25,000 cp/mL Beta-Globin: 75,000 cp/mL	Pass	10 Vials

**APPROVAL**

<b>Quality Assurance</b>	<b>Date</b>
	21 Jun 21