

PRODUCT INFORMATION

Product SKU	HPV18C
Product Name	HPV Genotype 18 Control
Kit Lot Number	21051107
Source Material	HPV Genotype 18 Cell Line
Date of Expiration	2024-04-30
Base Matrix	PreservCyt®
Fill Volume	5.0 mL
Storage Temperature	2-8°C
Regulatory Status	CE-IVD
Testing Platform	In-House HPV-18 Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21051107	HPV Genotype 18 Control	HPV-18: 25,000 cp/mL	Pass	10 Vials
		Beta-Globin: 75,000 cp/mL		

APPROVAL

Quality Assurance	Date
	24 May 21