

PRODUCT INFORMATION

| | |
|---------------------|------------------------|
| Product SKU | HPVNEG |
| Product Name | HPV Negative Control |
| Kit Lot Number | 22050903 |
| Source Material | HPV Negative Cell Line |
| Date of Expiration | 2025-04-30 |
| Base Matrix | PreservCyt® |
| Fill Volume | 5.0 mL |
| Storage Temperature | 2-8°C |
| Regulatory Status | CE-IVD |
| Testing Platform | In-House HPV Assay |


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

| Lot Number | Product Description | Target Concentration | Testing Results (Comparison of Target Value Against Reference Standard) | Quantity |
|------------|----------------------|---------------------------|---|----------|
| 22050903 | HPV Negative Control | Beta-Globin: 75,000 cp/mL | Pass | 10 Vials |

APPROVAL

| | |
|---|-------------|
| Quality Assurance | Date |
|  | 16 May 22 |