

PRODUCT INFORMATION

Product SKU	HSV2H102
Product Name	HSV-2 High Run Control
Kit Lot Number	19050806
Source Material	Inactivated HSV-2
Date of Expiration	2020-10-31
Base Matrix	Synthetic CSF Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-house HSV-2 Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Testing Results	Quantity
19050806	HSV-2 High Run Control	Pass (Comparison of target value against reference standard)	5 vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
<i>Saylor Saylor</i>	12 Jun 19