

**PRODUCT INFORMATION**

Product SKU	HSV2L101
Product Name	HSV-2 Low Run Control
Kit Lot Number	20061604
Source Material	Inactivated HSV-2
Date of Expiration	2022-06-30
Base Matrix	Synthetic CSF Matrix
Preservative	ProClin™ 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House HSV-2 Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
20061604	HSV-2 Low Run Control	5,000 cp/mL	Pass	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	10 Nov 20

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107  
For assistance and reordering, contact Exact Diagnostics at 817.989.9262