

PRODUCT INFORMATION

Product SKU	HSV2P100
Product Name	HSV-2 Verification Panel
Kit Lot Number	19041516
Source Material	Inactivated HSV-2
Date of Expiration	2020-09-30
Base Matrix	Synthetic CSF Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House HSV-2 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
19041511	HSV-2 Panel Member 1: 2E2 cp/mL	200 cp/mL	Pass
19041512	HSV-2 Panel Member 2: 2E3 cp/mL	2,000 cp/mL	Pass
19041513	HSV-2 Panel Member 3: 2E4 cp/mL	20,000 cp/mL	Pass
19041514	HSV-2 Panel Member 4: 2E5 cp/mL	200,000 cp/mL	Pass
19041515	HSV-2 Panel Member 5: 2E6 cp/mL	2,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	30 Apr 19

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com
HSV2P100 LN19041516 Rev 01.00