

**PRODUCT INFORMATION**

<b>Product SKU</b>	HSV2P100
<b>Product Name</b>	HSV-2 Verification Panel
<b>Kit Lot Number</b>	20022010
<b>Source Material</b>	Inactivated HSV-2
<b>Date of Expiration</b>	2022-01-31
<b>Base Matrix</b>	Synthetic CSF Matrix
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House HSV-2 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
20022005	HSV-2 Panel Member 1: 2E2 cp/mL	200 cp/mL	Pass
20022006	HSV-2 Panel Member 2: 2E3 cp/mL	2,000 cp/mL	Pass
20022007	HSV-2 Panel Member 3: 2E4 cp/mL	20,000 cp/mL	Pass
20022008	HSV-2 Panel Member 4: 2E5 cp/mL	200,000 cp/mL	Pass
20022009	HSV-2 Panel Member 5: 2E6 cp/mL	2,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	03 Mar 20

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com

HSV2P100 LN20022010 Rev 01.00