

PRODUCT INFORMATION

Product SKU	HV6AP100
Product Name	HHV-6A Verification Panel
Kit Lot Number	19102410
Source Material	Inactivated HHV-6A
Date of Expiration	2021-09-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House HHV-6A Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
19102405	HHV-6A Panel Member 1: 5E2 cp/mL	500 cp/mL	Pass
19102406	HHV-6A Panel Member 2: 5E3 cp/mL	5,000 cp/mL	Pass
19102407	HHV-6A Panel Member 3: 5E4 cp/mL	50,000 cp/mL	Pass
19102408	HHV-6A Panel Member 4: 5E5 cp/mL	500,000 cp/mL	Pass
19102409	HHV-6A Panel Member 5: 5E6 cp/mL	5,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
<i>Daylen Apeli</i>	06 Nov 19