

PRODUCT INFORMATION

Product SKU	HV6BP100
Product Name	HHV-6B Verification Panel
Kit Lot Number	22013101
Source Material	Inactivated HHV-6B
Date of Expiration	2024-01-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Traceability	1 st WHO International Standard (15/266)
Testing Platform	In-House HHV-6B Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations		Results
		Copies	IU	
22013102	HHV-6B Panel Member 1: 5E2 cp/mL	500 cp/mL	1,250 IU/mL	Pass
22013103	HHV-6B Panel Member 2: 5E3 cp/mL	5,000 cp/mL	12,500 IU/mL	Pass
22013104	HHV-6B Panel Member 3: 5E4 cp/mL	50,000 cp/mL	125,000 IU/mL	Pass
22013105	HHV-6B Panel Member 4: 5E5 cp/mL	500,000 cp/mL	1,250,000 IU/mL	Pass
22013106	HHV-6B Panel Member 5: 5E6 cp/mL	5,000,000 cp/mL	12,500,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	23Feb22