

PRODUCT INFORMATION

Product SKU	HV6H102
Product Name	HHV-6 High Run Control
Kit Lot Number	20103021
Source Material	Inactivated HHV-6B
Date of Expiration	2022-10-31
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Traceability	1 st WHO International Standard (15/266)
Testing Platform	In-House HHV-6B Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
20103021	HHV-6 High Run Control	100,000 cp/mL	Pass	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	30 NOV 20

Manufactured By: Exact Diagnostics, 3400 Camp Bowie Blvd. CBH-214, Fort Worth, Texas 76107

For assistance and reordering, contact Exact Diagnostics at 817.989.9262

www.exactdiagnostics.com

HV6H102 LN20103021 Rev 01.00