

**PRODUCT INFORMATION**

Product SKU	HV6L101
Product Name	HHV-6 Low Run Control
Kit Lot Number	21031205
Source Material	Inactivated HHV-6B
Date of Expiration	2023-02-28
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Traceability	1 <sup>st</sup> WHO International Standard (15/266)
Testing Platform	In-House HHV-6B Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21031205	HHV-6 Low Run Control	5,000 cp/mL	Pass	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	23 Mar 21