

PRODUCT INFORMATION

Product SKU	NEG000
Product Name	Negative Run Control
Kit Lot Number	20010203
Date of Expiration	2021-06-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	For In Vitro Diagnostics Use

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Testing Platform	Testing Results	Quantity
20010203	Negative Run Control	HIV-1	Roche cobas® TaqScreen MPX Test	Negative	5 Vials
		HBV	Roche cobas® TaqScreen MPX Test	Negative	
		HCV	Roche cobas® TaqScreen MPX Test	Negative	
		CMV	In-House CMV Test	Negative	
		EBV	In-House EBV Test	Negative	
		BKV	In-House BKV Test	Negative	
		Adenovirus	In-House Adenovirus Test	Negative	
		HHV-6	In-House HHV-6 Test	Negative	
		HSV-1	In-House HSV-1 Test	Negative	
		HSV-2	In-House HSV-2 Test	Negative	
		JCV	In-House JCV Test	Negative	
		VZV	In-House VZV Test	Negative	
		Zika Virus	In-House Zika Virus Test	Negative	
		HIV-2	In-House HIV-2 Test	Negative	

Bioburden Testing Results

Acceptance Criteria: Less than 10 CFU/mL	Pass
--	------

APPROVAL

Quality Assurance	Date
	16 Jun 20