

PRODUCT INFORMATION

Product SKU	NEG050
Product Name	Bulk Negative
Lot Number	21111712
Date of Expiration	2024-10-31
Date of Manufacture	2021-11-18
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	50.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Testing Platform	Testing Results
21111712	Bulk Negative	HIV-1	Roche cobas® TaqScreen MPX Test	Negative
		HBV	Roche cobas® TaqScreen MPX Test	Negative
		HCV	Roche cobas® TaqScreen MPX Test	Negative
		CMV	In-House CMV Test	Negative
		BKV	In-House BKV Test	Negative
		EBV	In-House EBV Test	Negative
		Adenovirus	In-House Adenovirus Test	Negative
		HHV-6	In-House HHV-6 Test	Negative
		HSV-1	In-House HSV-1 Test	Negative
		HSV-2	In-House HSV-2 Test	Negative
		JCV	In-House JCV Test	Negative
		VZV	In-House VZV Test	Negative
		Zika Virus	In-House Zika Virus Test	Negative

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance 	Date 30 Nov 21
---	--------------------------