

PRODUCT INFORMATION

Product SKU	RPNEG
Product Name	RP Negative Run Control
Lot Number	20050803
Date of Expiration	2021-10-31
Base Matrix	Stabilizing Matrix
Preservative	ProClin® 300
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House Respiratory Assays

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This QC material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.


TESTING DATA

Lot Number	Product Description	Analyte	Testing Results	Quantity
20050803	RP Negative Run Control	Adenovirus	Negative	6 Vials
		Bordetella parapertussis	Negative	
		Bordetella pertussis	Negative	
		Chlamydomydia pneumoniae	Negative	
		Coronavirus 229E	Negative	
		Coronavirus HKU	Negative	
		Coronavirus NL63	Negative	
		Coronavirus OC43	Negative	
		Human metapneumovirus	Negative	
		Influenza A H1N1	Negative	
		Influenza A H1N1-09 (RNA IVT)	Negative	
		Influenza A H3N2 (RNA IVT)	Negative	
		Influenza B	Negative	
		Mycoplasma pneumoniae	Negative	
		Parainfluenza 1	Negative	
		Parainfluenza 2	Negative	
		Parainfluenza 3	Negative	
		Parainfluenza 4a	Negative	
		Rhinovirus 1A	Negative	
RSV A	Negative			
RSV B	Negative			
MERS (RNA IVT)	Negative			

Continued next page

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	21 Jun 21