

PRODUCT INFORMATION

Product SKU	RPPOS
Product Name	RP Positive Run Control
Lot Number	19110404
Date of Expiration	2021-04-30
Base Matrix	Stabilizing Matrix
Preservative	ProClin® 300
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	BioFire FilmArray (RP & RP2)

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This QC material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.


TESTING DATA

Lot Number	Product Description	Analyte	Testing Results	Quantity
19110404	RP Positive Run Control	Adenovirus	Positive	6 Vials
		Bordetella parapertussis	Positive	
		Bordetella pertussis	Positive	
		Chlamydomphila pneumoniae	Positive	
		Coronavirus 229E	Positive	
		Coronavirus HKU	Positive	
		Coronavirus NL63	Positive	
		Coronavirus OC43	Positive	
		Human metapneumovirus	Positive	
		Influenza A H1N1-09	Positive	
		Influenza A H3N2	Positive	
		Influenza B	Positive	
		Mycoplasma pneumoniae	Positive	
		Parainfluenza 1	Positive	
		Parainfluenza 2	Positive	
		Parainfluenza 3	Positive	
		Parainfluenza 4a	Positive	
		Rhinovirus 1A	Positive	
RSV A	Positive			
RSV B	Positive			
MERS	Positive			

Continued next page

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	07 Jan 20