

**PRODUCT INFORMATION**

<b>Product SKU</b>	RPRC
<b>Product Name</b>	Respiratory Run Control
<b>Kit Lot Number</b>	18120415
<b>Source Material</b>	Inactivated Adenovirus
	Inactivated <i>Bordetella parapertussis</i>
	Inactivated <i>Bordetella pertussis</i>
	Inactivated <i>Chlamydophila pneumoniae</i>
	Inactivated Coronavirus 229E
	Inactivated Coronavirus OC43
	Inactivated Enterovirus
	Inactivated Influenza A H1N1
	Inactivated Influenza A H1N1-2009
	Inactivated Influenza A H3N2
	Inactivated Influenza B
	Inactivated <i>Mycoplasma pneumoniae</i>
	Inactivated Parainfluenza 1
	Inactivated Parainfluenza 2
	Inactivated Parainfluenza 3
	Inactivated Parainfluenza 4
<b>Date of Expiration</b>	2020-01-31
<b>Base Matrix</b>	Stabilizing Matrix
<b>Preservatives</b>	ProClin® 300
<b>Fill Volume</b>	0.3 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For In Vitro Diagnostic Use (IVD)
<b>Testing Platform</b>	In-House Assays and/or Luminex Verigene System

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Analyte	Testing Results
18120411	Respiratory Control Mix 1	Coronavirus 229E	Pass
		Influenza A H1N1	
		Influenza A H3N2	
		Influenza B	
		RSV A	
18120412	Respiratory Control Mix 2	Adenovirus	Pass
		<i>Bordetella parapertussis</i>	
		Enterovirus	
		Rhinovirus	
18120413	Respiratory Control Mix 3	Coronavirus OC43	Pass
		Parainfluenza 1	
		Parainfluenza 2	
		Parainfluenza 3	
		Parainfluenza 4	
18120414	Respiratory Control Mix 4	<i>Bordetella pertussis</i>	Pass
		<i>Chlamydomphila pneumoniae</i>	
		Influenza A H1N1-2009	
		<i>Mycoplasma pneumoniae</i>	
		RSV B	

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	11 Mar 19