

PRODUCT INFORMATION

Product SKU	COV000
Product Name	SARS-CoV-2 Negative
Lot Number	20030814
Date of Expiration	2021-02
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	For Research Use Only. Not for use in diagnostic procedures.
Testing Platform	CDC 2019-Novel Coronavirus (2019-nCoV) and Altona RealStar SARS-CoV-2 RT-PCR Kit 1.0


This standard material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Testing Results	Quantity
20030814	SARS-CoV-2 Negative	Envelope (E) gene	Negative	5 Vials
		Nucleocapsid (N) gene	Negative	
		Spike Protein (S) gene	Negative	
		Human gDNA	75,000 copies/mL	

APPROVAL

Quality Assurance	Date
	12 Mar 20