

PRODUCT INFORMATION

Product SKU	TVMGP200
Product Name	TV/MG Verification Panel
Kit Lot Number	22021113
Source Material 1	<i>Trichomonas vaginalis</i>
Source Material 2	<i>Mycoplasma genitalium</i>
Date of Expiration	2023-07-31
Base Matrix	Multiprep Specimen Diluent
Preservative	Sodium Azide
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House TV Assay
	In-House MG Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
22021104	TV 3x LOD	TV: 900 cp/mL	Pass
22021105	TV 5x LOD	TV: 1,500 cp/mL	Pass
22021106	TV 10x LOD	TV: 3,000 cp/mL	Pass
22021107	MG 3x LOD	MG: 900 cp/mL	Pass
22021108	MG 5x LOD	MG: 1,500 cp/mL	Pass
22021109	MG 10x LOD	MG: 3,000 cp/mL	Pass
22021110	TV/MG 3x LOD	TV: 900 cp/mL	Pass
		MG: 900 cp/mL	
22021111	TV/MG 5x LOD	TV: 1,500 cp/mL	Pass
		MG: 1,500 cp/mL	
22021112	TV/MG 10x LOD	TV: 3,000 cp/mL	Pass
		MG: 3,000 cp/mL	

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	01 Mar 22