

**PRODUCT INFORMATION**

<b>Product SKU</b>	TVMGP200
<b>Product Name</b>	TV/MG Verification Panel
<b>Kit Lot Number</b>	64514780
<b>Source Material 1</b>	<i>Trichomonas vaginalis</i>
<b>Source Material 2</b>	<i>Mycoplasma genitalium</i>
<b>Date of Expiration</b>	2024-03-31
<b>Base Matrix</b>	Multiprep Specimen Diluent
<b>Preservative</b>	Sodium Azide
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House TV Assay
	In-House MG Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
64514780-1	TV 3x LOD	TV: 900 cp/mL	Pass
64514780-2	TV 5x LOD	TV: 1,500 cp/mL	Pass
64514780-3	TV 10x LOD	TV: 3,000 cp/mL	Pass
64514780-4	MG 3x LOD	MG: 900 cp/mL	Pass
64514780-5	MG 5x LOD	MG: 1,500 cp/mL	Pass
64514780-6	MG 10x LOD	MG: 3,000 cp/mL	Pass
64514780-7	TV/MG 3x LOD	TV: 900 cp/mL	Pass
		MG: 900 cp/mL	
64514780-8	TV/MG 5x LOD	TV: 1,500 cp/mL	Pass
		MG: 1,500 cp/mL	
64514780-9	TV/MG 10x LOD	TV: 3,000 cp/mL	Pass
		MG: 3,000 cp/mL	

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
<i>[Signature]</i>	250422