

PRODUCT INFORMATION

Product SKU	TVMGP200
Product Name	TV/MG Verification Panel
Kit Lot Number	21072715
Source Material 1	<i>Trichomonas vaginalis</i>
Source Material 2	<i>Mycoplasma genitalium</i>
Date of Expiration	2023-01-31
Base Matrix	Multiprep Specimen Diluent
Preservative	Sodium Azide
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House TV Assay In-House MG Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Results
21072706	TV 3x LOD	Pass
21072707	TV 5x LOD	Pass
21072708	TV 10x LOD	Pass
21072709	MG 3x LOD	Pass
21072710	MG 5x LOD	Pass
21072711	MG 10x LOD	Pass
21072712	TV/MG 3x LOD	Pass
21072713	TV/MG 5x LOD	Pass
21072714	TV/MG 10x LOD	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance 	Date 13 Sept 21
---	---------------------------