

PRODUCT INFORMATION

Product SKU	VZVH102
Product Name	VZV High Run Control
Kit Lot Number	19100901
Source Material	Inactivated VZV
Date of Expiration	2021-09-30
Base Matrix	Synthetic CSF Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House VZV Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Testing Results	Quantity
19100901	VZV High Run Control	Pass (Comparison of target value against reference standard)	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
<i>Taylor Taylor</i>	23 Oct 19