

EDX CMV High Run Control Summary

Catalog Number	CMVH102
Analyte	Inactivated CMV
Matrix	True EDTA Plasma
Preservative	ProClin® 300
Storage	-20°C or Below
Fill Volume	1.0mL
Number of Vials Per Kit	5 Vials
Number of Uses	Multiple Use
Shelf Life	36 Months from Date of Manufacture
Traceability	1 st WHO International Standard
Precautions	Biological Risks
Regulatory Status	For In Vitro Diagnostic Use

INTENDED USE:

The EDX CMV High Run Control is an external quality control, intended to be used with CMV assays to monitor the performance of that given assay. EDX CMV High Run Control is calibrated against the 1st WHO International Standard for Cytomegalovirus (09/162)¹. The control is used to monitor the presence of cytomegalovirus (CMV) DNA. Routine use of EDX CMV High Run Control allows laboratories to evaluate the day-to-day and lot-to-lot variation of their molecular assay and test for operator proficiency.

PRODUCT DESCRIPTION:

The EDX CMV High Run Control contains whole, intact virus that have been heat inactivated. The control is intended to test the entire process of molecular assays including extraction, detection, and amplification.

EDX CMV High Run Control is formulated in True EDTA Plasma that is non-reactive for HBV DNA, HCV RNA, HIV-1 RNA, CMV DNA, HBsAg and antibodies for HIV-1/2, and HCV.

PROCEDURE:

The EDX CMV High Run Control requires an extraction step. The control must be treated in a similar manner to other tested samples. This allows the operator to monitor their extraction technique.

The EDX CMV High Run Control is an unassayed control and does not have assigned values. The control is not to be used as a substitute for the manufacturer's kit controls provided with the assay. EDX CMV High Run Control is to be tested according to the assay manufacturer's or testing laboratory's instructions and recommendations. Expected results utilizing EDX CMV High Run Control must be established by the end user for their specific assay.

STORAGE AND HANDLING:

The EDX CMV High Run Control should be stored at -20°C or below.

Thaw EDX CMV High Run Control at room temperature and vortex briefly prior to use.

Once thawed, EDX CMV High Run Control is stable for 24 hours when stored at 2-8°C. Product can be frozen and reused a maximum of four (4) times. After the fourth freeze thaw, any remaining materials must be disposed of. Do not dilute.

Do not use EDX CMV High Run Control beyond the expiration date.

LIMITATIONS:

For In Vitro Diagnostic Use.

WARNINGS AND PRECAUTIONS:

The EDX CMV High Run Control contains heat inactivated virus but should be considered biohazardous. Universal precautions and proper disposal should be practiced². Do not pipette by mouth. Do not smoke, eat, or drink in areas where specimens are handled.

Discard product if packaging is damaged or leaking. Disinfect liquids, materials or spills with a 0.5% sodium hypochlorite solution. Dispose of all materials and liquids used in the procedure as if they contained pathogenic agents.

REFERENCES:

¹The 1st WHO International Standard for Cytomegalovirus (09/162) IFU. <http://www.nibsc.org/documents/ifu/09-162.pdf>

² Siegel JD, Rhinehart E, Jackson M, Chiarello L, and the Healthcare Infection Control Practices Advisory Committee, 2007 Guideline for Isolation Precautions: Preventing Transmission

SYMBOL REFERENCES:



Catalog Number



Lot Number



For In Vitro Diagnostic Use



Biological Risks



Expiration Date



Upper Limit of Temperature



Manufacturer



Caution



Consult Instructions for Use




Positive Control



European Mark of Conformity



Authorized Representative



Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands



Exact Diagnostics LLC
100 South Jones Street, Suite 100,
Fort Worth, Texas 76104
United States



**Bio-Rad
Laboratories**

**Clinical
Diagnostics Group**

9500 Jeronimo Road
Irvine, California 92618
(800) 854-6737
FAX (949) 598-1550
bio-rad.com/qualitycontrol

Technical Service:
(800) 854-6737

United States Customer Service: 4000 Alfred Nobel Drive • Hercules, California 94547 • Phone (510) 724-7000 • FAX (510) 741-6373
Also in: **Gladesville, Australia,** Phone +61 (2) 9914 2800 • Fax +61 (2) 9914 2888 **Vienna, Austria,** Phone +43 (0) 1 877 89 01 9 • Fax +43 (0) 1 876 56 29 **Temse, Belgium,** Phone +32 (0) 3 710 53 00 • Fax +32 (0) 3 710 53 01 **São Paulo, Brazil,** Phone +55 11 3065-7550 **Montréal, Canada,** Phone +1 514 334 4372 • Fax +1 514 334 0872 **Shanghai, China,** Phone +86 21 6169 8500 • Fax +86 21 6169 8599 **Prague, Czech Republic,** Phone +420 241 431 660 **Symbion Science Park, Denmark,** Phone +45 44 52 10 00 • Fax +45 44 52 10 01 **Helsinki, Finland,** Phone +358 9 804 22 00 **Marnes-la-Coquette, France,** Phone +33 (0)1 47 95 60 00 • Fax +33 (0)1 47 41 91 33 **Munich, Germany,** Phone +49 (0) 89 31804 393 • Fax +49 (0) 89 31804 136 **Athens, Greece,** Phone +30 210 7774396 • Fax +30 210 7774376 **Quarry Bay, Hong Kong,** Phone +85 2 2789 3300 • Fax +85 2 2789 1257 **Budapest, Hungary,** Phone +36 1 459 6190 • Fax +36 1 459 6101 **Haryana, India,** Phone +91 124 4029300 • Fax +91 124 2398115 **Rishon Le Zion, Israel,** Phone +972 03 963 6025 • Fax +972 03 951 4129 **Milan, Italy,** Phone +39 024 94 86 600 • Fax +39 02 21609399 **Tokyo, Japan,** Phone +81 3 6361 7070 • Fax +81 3 5463 8481 **Seoul, Korea,** Phone +82 080 007 7373 • Fax +82 (2) 3472 7003 **Mexico D.F., Mexico,** Phone +52 (55) 5488 7670 • Fax +52 (55) 1107 7246 **Veenendaal, The Netherlands,** Phone +31 (0)318 540 666 • Fax +31 (0)318 542 216 **Auckland, New Zealand,** Phone +64 (9)415 2280 • Fax +64 (9)415 2284 **Oslo, Norway,** Phone +47 23 38 41 30 • Fax +47 23 38 41 39 **Warsaw, Poland,** Phone +48 22 331 99 99 • Fax +48 22 331 99 88 **Amadora, Portugal,** Phone +351 21 47 27 700 • Fax +351 21 472 7777 **Moscow, Russia,** Phone +7 495 721 1404 • Fax +7 495 721 1412 **Singapore,** Phone +65 6415 3170 • Fax +65 6415 3189 **Johannesburg, South Africa,** Phone +27 11 442 8508 • Fax +27 11 442 8525 **Madrid, Spain,** Phone +34 91 490 6580 • Fax +34 91 590 5211 **Sundbyberg, Sweden,** Phone +46 844 98053 • Fax +46 8 55 51 27 80 **Cressier, Switzerland,** Phone +41 (0) 61 717 9555 • Fax +41 (0) 61 717 9550 **Taipei, Taiwan,** Phone +886 (2) 2578-7189 • Fax +886 (2) 2578-6890 **Bangkok, Thailand,** Phone (662) 651 8311 • Fax (662) 651 8312 **Hertfordshire, United Kingdom,** Phone +44 (0)1923 471301 • Fax +44 (0)1923 471340