

PRODUCT INFORMATION

Product SKU	COV000CE
Product Name	SARS-CoV-2 Negative Run Control
Lot Number	64550504
Date of Manufacture	2023-05-01
Date of Expiration	2024-10-31
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	0.3 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House RNase P Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Analyte	Target Concentration	Testing Results	Quantity
64550504	SARS-CoV-2 Negative Run Control	Human gDNA	75,000 cp/mL RNase P DNA	Pass 251,900 cp/mL RNase P (DNA + RNA)	5 Vials

APPROVAL

Quality Assurance	Date
<i>Melody Hays</i>	<i>11 May 23</i>