

**PRODUCT INFORMATION**

<b>Product SKU</b>	COVFLU
<b>Product Name</b>	SARS-CoV-2, Flu, RSV Positive Run Control
<b>Lot Number</b>	21121001
<b>Date of Expiration</b>	2023-11-30
<b>Base Matrix</b>	Stabilizing Matrix
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House Influenza A H1N1, Influenza B, RSV A, N Gene, RNase P Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Analyte	Target Concentrations	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
21121001	SARS-CoV-2, Flu, RSV Positive Run Control	Influenza A H1N1	10,000 cp/mL	Pass	5 Vials
		Influenza B	10,000 cp/mL	Pass	
		RSV A	10,000 cp/mL	Pass	
		SARS-CoV-2	26,300 cp/mL	Pass	
		Human gDNA	10,000 cp/mL RNase P DNA	Pass	

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

<b>Quality Assurance</b>	<b>Date</b>
	18 Jan 22