

**PRODUCT INFORMATION**

<b>Product SKU</b>	COVFLUP200
<b>Product Name</b>	SARS-CoV-2, Flu, RSV Verification Panel
<b>Kit Lot Number</b>	64528221
<b>Source Material 1</b>	Inactivated Influenza A - H1N1
<b>Source Material 2</b>	Inactivated Influenza B - B/Lee/40
<b>Source Material 3</b>	Inactivated SARS-CoV-2 - 2019-nCoV/USA-WA1/2020
<b>Source Material 4</b>	Inactivated RSV A - VR26 Long
<b>Date of Expiration</b>	2024-06-30
<b>Base Matrix</b>	Multiprep Specimen Diluent
<b>Fill Volume</b>	1.1 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	For Research Use Only. Not for use in diagnostic procedures.
<b>Testing Platform</b>	Roche cobas® SARS-CoV-2 & Influenza A/B assay on Roche cobas® 6800/8800 platform In-house RSV A assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Quantity	Results
64528221-1	SARS-CoV-2 3x LOD	1,800 cp/mL	1 Vial	Pass
64528221-2	SARS-CoV-2 5x LOD	3,000 cp/mL	1 Vial	Pass
64528221-3	SARS-CoV-2 10x LOD	6,000 cp/mL	1 Vial	Pass
64528221-4	Flu A 3x LOD	900 cp/mL	1 Vial	Pass
64528221-5	Flu A 5x LOD	1,500 cp/mL	1 Vial	Pass
64528221-6	Flu A 10x LOD	3,000 cp/mL	1 Vial	Pass
64528221-7	Flu B 3x LOD	900 cp/mL	1 Vial	Pass
64528221-8	Flu B 5x LOD	1,500 cp/mL	1 Vial	Pass
64528221-9	Flu B 10x LOD	3,000 cp/mL	1 Vial	Pass
64528221-10	RSV A 3x LOD	2,400 cp/mL	1 Vial	Pass
64528221-11	RSV A 5x LOD	4,000 cp/mL	1 Vial	Pass
64528221-12	RSV A 10x LOD	8,000 cp/mL	1 Vial	Pass
64528221-13	SARS-CoV-2, Flu, RSV 3x LOD	SARS-CoV-2: 1,800 cp/mL	1 Vials	Pass
		Flu A: 900 cp/mL		

		Flu B: 900 cp/mL		
		RSV A: 2,400 cp/mL		
64528221-14	SARS-CoV-2, Flu, RSV 5x LOD	SARS-CoV-2: 3,000 cp/mL	1 Vial	Pass
		Flu A: 1,500 cp/mL		
		Flu B: 1,500 cp/mL		
		RSV A: 4,000 cp/mL		
64528221-15	SARS-CoV-2, Flu, RSV 10x LOD	SARS-CoV-2: 6,000 cp/mL	1 Vial	Pass
		Flu A: 3,000 cp/mL		
		Flu B: 3,000 cp/mL		
		RSV A: 8,000 cp/mL		

**APPROVAL**

Quality Assurance	Date
<i>Melody Hantz</i>	<i>17 FEB 23</i>